



THE CUBE

for Laboratory Applications


Ambient Air Gas Detector



The Otis Instruments OI-6000 Series Ambient Air Toxic Gas Detector provides ease of use and implementation to comply with varying demands within the laboratory industry.

The OI-6000 is a versatile assembly that uses an electrochemical, infrared or PID sensor element to detect a variety of gases. Wired power (12 to 35 VDC) with 4-20 mA analog and RS-485 Modbus outputs. Magnetic switches allow for non-intrusive calibration while in the field as well as full system settings and diagnostics.

OTIS
INSTRUMENTS INC.
www.otisinstruments.com



THE CUBE

Product Specifications	
Sensor Type	Electrochemical, Infrared, Photo Ionization Detector
Power Voltage	Wired Power (2-Wire) 12 to 35 VDC (Optional AC Power Supply)
Power Draw	2 Watts Maximum
Protection	Power Electromagnetic Interference (EMI) Filter 4-20 mA Surge Suppression RS-485 Surge Suppression
Display Screen	102x64 Resolution Graphical LCD Screen Transflective (Sunlight Readable) with White LED Backlight
Interface	3 Push-Buttons (MENU, ADD and SUB) 3 Magnetic Non-Intrusive Switches for Calibration (MENU, ADD and SUB)
Customizable Settings	2 Individually-Configurable Alarm LED Set-Points
Tamb Temperature Range	-20°C to +54°C / -4°F to 130°F
Wired (Analog) Output	4-20 mA (3-Wire) - Fault Terminal
(Digital) Output	RS-485 Modbus Optional GEN II 900 MHz or Gen II 2.4 GHz Radio WireFree Output
Product Dimensions	5" L x 5" W x 5" H
Total Weight	5 lbs.
Hardware Warranty	Limited one year
Sensor Element Warranty	Varies with Gas Type
Part Number Formula	OI-6000-[Sensor Type]-[Gas]-[Radio Option]-[Relay Option]-C

Gas Type	Range	T ₉₀	Warranty
CL2	0-10 ppm	< 60 sec.	12 months
CLO2	0-1 ppm	< 120 sec.	12 months
CO	0-1000 ppm	< 30 sec.	24 months
CO2	0-5% vol.	< 30 sec.	60 months
H2	0-4% vol.	< 60 sec.	18 months
H2S	0-100 ppm	< 30 sec.	24 months
H2S2K	0-2000 ppm	< 25 sec.	24 months
HCL	0-30 ppm	< 300 sec.	12 months
HCN	0-50 ppm	< 70 sec.	12 months
LEL	0-100% (0-5 % vol.)	< 30 sec.	60 months
NH3	0-100 ppm	< 90 sec.	12 months
NH3A	0-1000 ppm	< 120 sec.	12 months
NO2	0-20 ppm	< 25 sec.	12 months
O2	0-25% vol.	< 15 sec.	24 months
PH3	0-5 ppm	< 30 sec.	12 months
SO2	0-20 ppm	< 35 sec.	24 months
VOC	0-2000 ppm (Max)	< 3 sec.	6 months

